

## Hastings Prince Edward Public Health Infection Prevention and Control Lapse Report

<b>Initial Report</b>				
Premise/Facility Under Investigation Dr. Ademidun- Obstetrics and Gynecology				
<b>Address</b>				
Unit Number 203	Street Number 210		Street Name Dundas	
City/Town Belleville	Province Ontario		Postal Code K8N 5G8	
Type of Premises/Facility Obstetrics and Gynecology				
Date Board of Health became aware of Potential IPAC Lapse 04/06/2021			Date of Initial Report Posting 18/06/2021	
Date of Initial report Update(s) (if applicable) <a href="#">Click here to enter a date.</a>			How the Potential IPAC Lapse was Identified Complaint	
<b>Summary Description of the IPAC Lapse</b> During the complaint inspection it was documented that the cleaning, disinfection and sterilization of medical equipment and devices on-site did not follow the Provincial Infectious Diseases Advisory Committee (PIDAC) Best Practice Standards for Medical Device Reprocessing and Infection Control in the Clinical Office Setting. Biological indicators were not being used in the reprocessing of medical instruments and the sterilizer was not properly tested prior to use as per the Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings. Sterilization records were not available at the time of the inspection.				
<b>IPAC Lapse Investigation</b>	<b>Yes</b>	<b>No</b>	<b>N/A</b>	<b>Please provide further details/steps</b>
Did the IPAC lapse involve a member of a regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	College of Physicians and Surgeons of Ontario (CPSO)
If yes, was the issue referred to the regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	CPSO was notified of IPAC lapse via e-mail.
Were any corrective measures recommended and/or implemented?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Single use medical equipment/devices were being used as per Hastings Prince Edward Public Health's Order.
Date any order(s) or Directives (s) were issued to the owner/operator (if applicable) Verbal Order issued 2021/06/09 Written Order issued 2021/06/11				
<b>Initial Report Comments</b>				
Ensure that only single use medical devices are used for medical procedures until the sterilizer has passed the appropriate challenges (i.e., biological, chemical) in accordance with the Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings. Staff members responsible for any or all steps in reprocessing must complete Public Health Ontario's (PHO) reprocessing learning modules before reprocessing medical equipment. Proper documentation of cleaning, disinfecting and sterilization must be recorded in accordance with PIDAC's Best Practice Standards for medical device reprocessing and infection control in the clinical office setting.				
<b>Final Report</b>				
Date of Final Report Posting 28/06/2021				
Date of all corrective measures were confirmed 25/06/2021				
Brief description of corrective measures taken Single use equipment/devices were being used for medical procedures/examinations done in the clinic while HPEPH's Order was in effect. The sterilizer has since passed the appropriate challenges (i.e., biological, chemical) and records are being maintained. Staff responsible for reprocessing have completed PHO's reprocessing learning modules.				

## Hastings Prince Edward Public Health Infection Prevention and Control Lapse Report

<b>Final Report Comments and Contact Information</b>		
<p><b>Any Additional Comments:</b></p> <p>Dr. Ademidun's obstetrics and gynecology office at 210 Dundas St. E. in Belleville, ON. is now in compliance with the reprocessing requirements and recommendations made by Hastings Prince Edward Public Health as per the Provincial Infectious Diseases Advisory Committee (PIDAC) Best Practice Standards for Medical Device Reprocessing and Infection Control in the Clinical Office Setting and the Best Practices for Cleaning, Disinfection and Sterilization of Medical Equipment/Devices in All Health Care Settings.</p>		
If you have any further questions, please contact: Brittany Peters		
<p><b>Name</b></p> <p>Brittany Peters</p>	<p><b>Telephone Number</b></p> <p>613-966-5500 ext. 503</p>	<p><b>Email Address</b></p> <p>bpeters@hpeph.ca</p>